



Piraeus, 14/01/2026

Clarifications (Part 1)

Q1: Can you please clarify, if all the requested documents, economical offers, Technical offers, etc., should be in English or if they can be submitted in Greek?

A1: The documents can be submitted in Greek.

Q2: Regarding the Sediment Sampler, what must be the length of the Deployment Wire and also what does it meant by "Deck Hardware"?

A2: The deployment wire Indicatively has to cover depths of up to 30m.
Deck Hardware any available lifting & hoisting equipment that can be placed on the pier for operating the grab